

<b>1.</b> Device Name	<b>STRETCH PF</b>																																			
<b>2.</b> Description	Powder Free Vinyl, Beaded Cuff, Ambidextrous, Non Sterile, Does not contain DOP, DEHP, DBP, BBP, DNOP or DIDP phthalates, Single Use.																																			
<b>3.</b> Product Codes & Sizes	<table border="1"> <thead> <tr> <th>Product Code</th> <th>Size</th> <th>Inner Barcode</th> <th>Shipper Barcode</th> </tr> </thead> <tbody> <tr> <td>41025</td> <td><b>SMALL</b></td> <td>9311059410254</td> <td>19311059410251</td> </tr> <tr> <td>41026</td> <td><b>MEDIUM</b></td> <td>9311059410261</td> <td>19311059410268</td> </tr> <tr> <td>41027</td> <td><b>LARGE</b></td> <td>9311059410278</td> <td>19311059410275</td> </tr> <tr> <td>41028</td> <td><b>XL</b></td> <td>9311059410285</td> <td>19311059410282</td> </tr> </tbody> </table>				Product Code	Size	Inner Barcode	Shipper Barcode	41025	<b>SMALL</b>	9311059410254	19311059410251	41026	<b>MEDIUM</b>	9311059410261	19311059410268	41027	<b>LARGE</b>	9311059410278	19311059410275	41028	<b>XL</b>	9311059410285	19311059410282												
Product Code	Size	Inner Barcode	Shipper Barcode																																	
41025	<b>SMALL</b>	9311059410254	19311059410251																																	
41026	<b>MEDIUM</b>	9311059410261	19311059410268																																	
41027	<b>LARGE</b>	9311059410278	19311059410275																																	
41028	<b>XL</b>	9311059410285	19311059410282																																	
<b>4.</b> Recommended Applications	<ul style="list-style-type: none"> <li>Medical</li> <li>Dental</li> <li>Hospitals</li> <li>Diagnostic</li> <li>Pharmaceutical</li> <li>Laboratory</li> <li>Childcare</li> <li>Aged Care</li> <li>Emergency Services</li> <li>Food Processing</li> <li>Food Service</li> <li>Industrial</li> <li>Cleaning</li> <li>Cosmetic</li> </ul>																																			
<b>5.</b> Materials	Polyvinyl Chloride, Di-octyl Adipate, Dinonyl phthalate, Pentanediol Diisobutyrate, Epoxidised Soybean Oil, PU release agent, Calcium-Zinc Stabilizer, Stan-Tone HCC-29556																																			
<b>6.</b> Production Process	Raw material inspection, PVC mixing and preparation, Laboratory test of compounded PVC solution, Dip Line process, QC in process inspection, Tumbling, QC off-line inspection, Packing, Packing process inspection, QC Pre-shipment inspection, Ship.																																			
<b>7.</b> Standards & Approvals	<div style="display: flex; align-items: center;"> <div style="margin-right: 20px;"> </div> <div style="margin-right: 20px;"> </div> <div> <p><b>Stretch PF</b> is tested to and complies with;</p> <ul style="list-style-type: none"> <li>EN455 1-2:2009</li> <li>Entered into the ARTG - ARTG# 114520</li> <li>Acceptable Quality Level AQL 1.5</li> </ul> </div> </div>																																			
<b>8.</b> Dimensions	<table border="1"> <thead> <tr> <th>SIZE</th> <th>WEIGHT p/pc +0.3- 0.3gm</th> <th>LENGTH (mm)</th> <th>THICKNESS <b>FINGER</b> SINGLE WALL (min)</th> <th>THICKNESS <b>PALM</b> SINGLE WALL (min)</th> <th>PALM WIDTH (+/-5)</th> </tr> </thead> <tbody> <tr> <td><b>SMALL</b></td> <td>4.5gm</td> <td>240mm</td> <td>0.07mm</td> <td>0.07mm</td> <td>85mm</td> </tr> <tr> <td><b>MEDIUM</b></td> <td>5.0gm</td> <td>240mm</td> <td>0.07mm</td> <td>0.07mm</td> <td>95mm</td> </tr> <tr> <td><b>LARGE</b></td> <td>5.5gm</td> <td>240mm</td> <td>0.07mm</td> <td>0.07mm</td> <td>105mm</td> </tr> <tr> <td><b>XL</b></td> <td>6.0gm</td> <td>240mm</td> <td>0.07mm</td> <td>0.07mm</td> <td>115mm</td> </tr> </tbody> </table> <p><i>Note: Cuff Thickness reference only</i></p>						SIZE	WEIGHT p/pc +0.3- 0.3gm	LENGTH (mm)	THICKNESS <b>FINGER</b> SINGLE WALL (min)	THICKNESS <b>PALM</b> SINGLE WALL (min)	PALM WIDTH (+/-5)	<b>SMALL</b>	4.5gm	240mm	0.07mm	0.07mm	85mm	<b>MEDIUM</b>	5.0gm	240mm	0.07mm	0.07mm	95mm	<b>LARGE</b>	5.5gm	240mm	0.07mm	0.07mm	105mm	<b>XL</b>	6.0gm	240mm	0.07mm	0.07mm	115mm
SIZE	WEIGHT p/pc +0.3- 0.3gm	LENGTH (mm)	THICKNESS <b>FINGER</b> SINGLE WALL (min)	THICKNESS <b>PALM</b> SINGLE WALL (min)	PALM WIDTH (+/-5)																															
<b>SMALL</b>	4.5gm	240mm	0.07mm	0.07mm	85mm																															
<b>MEDIUM</b>	5.0gm	240mm	0.07mm	0.07mm	95mm																															
<b>LARGE</b>	5.5gm	240mm	0.07mm	0.07mm	105mm																															
<b>XL</b>	6.0gm	240mm	0.07mm	0.07mm	115mm																															
<b>9.</b> Physical Properties	<table border="1"> <thead> <tr> <th>DESCRIPTION</th> <th>BEFORE AGEING</th> <th>AFTER AGEING</th> </tr> </thead> <tbody> <tr> <td>Tensile Strength (N)</td> <td>3.6 minimum</td> <td>3.6 minimum</td> </tr> </tbody> </table>						DESCRIPTION	BEFORE AGEING	AFTER AGEING	Tensile Strength (N)	3.6 minimum	3.6 minimum																								
DESCRIPTION	BEFORE AGEING	AFTER AGEING																																		
Tensile Strength (N)	3.6 minimum	3.6 minimum																																		
<b>10.</b> Colour	Yellow / Cream																																			



**11. Performance Requirements**

Gloves shall be sampled and inspected in accordance with ISO 2859-1. The inspection levels and acceptable quality limits (AQLs) shall conform to those listed in below table for the characteristics listed.

CHARACTERISTICS	INSPECTION LEVEL	ACCEPTABLE QUALITY LEVEL	REFERENCE STANDARD
Dimensions	S2	4.0	EN455-2:2009
Physical Properties	S2	4.0	EN455-2:2009
Watertightness	G1	1.5	EN455-1:2009
Tensile Strength and Elongation at break (before and after accelerated ageing)	S2	4.0	EN455-2:2009
Visual Defects:			Manufacturers Practice
Critical	G1	1.5	
Major	G1	2.5	
Minor	G1	4.0	

**12. Quality Control Inspection Procedures**

- a) Shipment is placed in quarantine area separately from other gloves in warehouse as one lot
- b) Shipment is then sorted by sizes if shipment consists of more than one size.
- c) QC inspector records lot number(s) indicted on inner boxes and outer cases, sizes quantity of shipment, buyer, date, etc
- d) QC inspector selects at random samples according to ISO procedures for sampling, each lots of different sizes are handled separately
- e) Samples are sent for testing as per requirements
- f) Results of tests will be recorded and only those that meet acceptance criteria are released for dispatch. If shipment consists of more than one size, shipment will be released only when all lots of different sizes have passed.
- g) If a shipment is found to fail test requirements it will be rejected by QC inspector and not released for shipment.

**13. Packaging**

100 pieces will be packed into a dispenser box. The number contained in each dispenser box determined by weight. 10 dispenser boxes are packed in 2 layers of 5 giving 10 dispenser boxes inside export quality shipper carton (outer).

**14. Shipper Carton Specifications**

SIZE	LENGTH (mm)	WIDTH (mm)	HEIGHT (mm)	GROSS WEIGHT (kg)
SMALL	375	265	265	5.5
MEDIUM	375	265	265	6.5
LARGE	375	265	265	7.5
XL	375	265	265	8.0

**15. Storage Conditions**

The gloves will maintain their properties when stored in dry conditions out of direct sunlight at between 10°C to 30°C.

**16. Shelf Life**

The gloves shall have a shelf life of 3 years from the date of manufacture when stored in compliance with the above storage conditions.