

1. Device Name	SUPERSOFT																																								
2. Description	Powder Free Nitrile, Beaded Cuff, Textured Fingers, Ambidextrous, Non Sterile, Touchscreen Friendly and Silicone Free.																																								
3. Product Codes & Sizes	<table border="1"> <thead> <tr> <th>Product Code</th> <th>Size</th> <th>Inner Barcode</th> <th>Shipper Barcode</th> </tr> </thead> <tbody> <tr> <td>41130</td> <td>SMALL</td> <td>9311059411305</td> <td>19311059411302</td> </tr> <tr> <td>41131</td> <td>MEDIUM</td> <td>9311059411312</td> <td>19311059411319</td> </tr> <tr> <td>41132</td> <td>LARGE</td> <td>9311059411329</td> <td>19311059411326</td> </tr> <tr> <td>41133</td> <td>XL</td> <td>9311059411336</td> <td>19311059411333</td> </tr> </tbody> </table>				Product Code	Size	Inner Barcode	Shipper Barcode	41130	SMALL	9311059411305	19311059411302	41131	MEDIUM	9311059411312	19311059411319	41132	LARGE	9311059411329	19311059411326	41133	XL	9311059411336	19311059411333																	
Product Code	Size	Inner Barcode	Shipper Barcode																																						
41130	SMALL	9311059411305	19311059411302																																						
41131	MEDIUM	9311059411312	19311059411319																																						
41132	LARGE	9311059411329	19311059411326																																						
41133	XL	9311059411336	19311059411333																																						
4. Recommended Applications	<ul style="list-style-type: none"> Medical Medical Instrument Cleaning Hospitals Diagnostic Pharmaceutical Laboratory Dental Aged Care Abattoirs & Meatworks Correctional Forensic Food Processing Food Service Tattooists Emergency Services Cosmetic Veterinary Childcare 																																								
5. Materials	Nitrile -butadiene-rubber (NBR), Sulphur, Zinc Oxide, Titanium Dioxide, Antioxidant ZDBC, ZDEC.																																								
6. Production Process	Former cleaning and washing, coagulant dipping, coagulant drying, nitrile dipping, nitrile drying oven, beading, leaching tank dip, main curing oven, cooling tank, chlorination tank, soak rinse tank, neutralizer, post leaching tank, blower, stripping, QC inspection, tumbling, packing, Pre-shipment inspection.																																								
7. Standards & Approvals	<div style="display: flex; align-items: center;"> <div style="margin-right: 20px;"> </div> <div style="margin-right: 20px;"> </div> <div> <p>SuperSoft is tested to and complies with;</p> <ul style="list-style-type: none"> EN ISO 14971:2019 + A11:2021 ASTM 6319 EN455 Certified under HACCP Food Safety Accreditation Entered into the ARTG - ARTG# 520227 Acceptable Quality Level AQL 1.5 </div> </div>																																								
8. Dimensions	<table border="1"> <thead> <tr> <th>SIZE</th> <th>LENGTH (-10mm)</th> <th>THICKNESS FINGER SINGLE WALL (min)</th> <th>THICKNESS PALM SINGLE WALL (min/max)</th> <th>THICKNESS CUFF SINGLE WALL (min/max)</th> <th>WEIGHT PER PIECE +0.1/-0.1gm</th> <th>PALM WIDTH +/-5mm</th> </tr> </thead> <tbody> <tr> <td>SMALL</td> <td>240mm</td> <td>0.06mm</td> <td>.05/.07mm</td> <td>.04/.06mm</td> <td>2.7gm</td> <td>85mm</td> </tr> <tr> <td>MEDIUM</td> <td>240mm</td> <td>0.06mm</td> <td>.05/.07mm</td> <td>.04/.06mm</td> <td>3.0gm</td> <td>95mm</td> </tr> <tr> <td>LARGE</td> <td>240mm</td> <td>0.06mm</td> <td>.05/.07mm</td> <td>.04/.06mm</td> <td>3.3gm</td> <td>105mm</td> </tr> <tr> <td>XL</td> <td>240mm</td> <td>0.06mm</td> <td>.05/.07mm</td> <td>.04/.06mm</td> <td>3.6gm</td> <td>115mm</td> </tr> </tbody> </table>						SIZE	LENGTH (-10mm)	THICKNESS FINGER SINGLE WALL (min)	THICKNESS PALM SINGLE WALL (min/max)	THICKNESS CUFF SINGLE WALL (min/max)	WEIGHT PER PIECE +0.1/-0.1gm	PALM WIDTH +/-5mm	SMALL	240mm	0.06mm	.05/.07mm	.04/.06mm	2.7gm	85mm	MEDIUM	240mm	0.06mm	.05/.07mm	.04/.06mm	3.0gm	95mm	LARGE	240mm	0.06mm	.05/.07mm	.04/.06mm	3.3gm	105mm	XL	240mm	0.06mm	.05/.07mm	.04/.06mm	3.6gm	115mm
SIZE	LENGTH (-10mm)	THICKNESS FINGER SINGLE WALL (min)	THICKNESS PALM SINGLE WALL (min/max)	THICKNESS CUFF SINGLE WALL (min/max)	WEIGHT PER PIECE +0.1/-0.1gm	PALM WIDTH +/-5mm																																			
SMALL	240mm	0.06mm	.05/.07mm	.04/.06mm	2.7gm	85mm																																			
MEDIUM	240mm	0.06mm	.05/.07mm	.04/.06mm	3.0gm	95mm																																			
LARGE	240mm	0.06mm	.05/.07mm	.04/.06mm	3.3gm	105mm																																			
XL	240mm	0.06mm	.05/.07mm	.04/.06mm	3.6gm	115mm																																			
9. Physical Properties	<table border="1"> <thead> <tr> <th>DESCRIPTION</th> <th>BEFORE AGEING</th> <th>AFTER AGEING</th> </tr> </thead> <tbody> <tr> <td>Elongation at break (%)</td> <td>500 minimum</td> <td>400 minimum</td> </tr> <tr> <td>Tensile Strength (MPa)</td> <td>14 minimum</td> <td>14 minimum</td> </tr> </tbody> </table> <p><i>Note: if Textured gloves then minimum requirements shall be 10% lower than those shown in table.</i></p>						DESCRIPTION	BEFORE AGEING	AFTER AGEING	Elongation at break (%)	500 minimum	400 minimum	Tensile Strength (MPa)	14 minimum	14 minimum																										
DESCRIPTION	BEFORE AGEING	AFTER AGEING																																							
Elongation at break (%)	500 minimum	400 minimum																																							
Tensile Strength (MPa)	14 minimum	14 minimum																																							
10. Colour	Blue																																								



11. Performance Requirements

Gloves shall be sampled and inspected in accordance with ISO 2859-1. The inspection levels and acceptable quality limits (AQLs) shall conform to those listed in below table for the characteristics listed.

CHARACTERISTICS	INSPECTION LEVEL	ACCEPTABLE QUALITY LEVEL	REFERENCE STANDARD
Dimensions	S2	4.0	ASTM 6319
Physical Properties	S2	4.0	ASTM 6319
Watertightness	G1	1.5	ASTM 6319
Tensile Strength and Elongation at break (before and after accelerated ageing)	S2	4.0	ASTM 6319
Visual Defects: Major	G1	2.5	Manufacturers Practice
Minor		2.5	
Packaging Defects Regulatory	G1	**	In House Inspection
Visual	G1	4.0	
Critical (Incl Glove Count)	S2	4.0	

Notes: ** indicates unacceptable at any level For Dimensions and Physical Properties a median of 13 test pieces used as per EN455-02-00

12. Quality Control Inspection Procedures

- a) Shipment is placed in quarantine area separately from other gloves in warehouse as one lot
- b) Shipment is then sorted by sizes if shipment consists of more than one size.
- c) QC inspector records lot number(s) indicated on inner boxes and outer cases, sizes quantity of shipment, buyer, date, etc
- d) QC inspector selects at random samples according to ISO procedures for sampling, each lots of different sizes are handled separately
- e) Samples are sent for testing as per requirements
- f) Results of tests will be recorded and only those that meet acceptance criteria are released for dispatch. If shipment consists of more than one size, shipment will be released only when all lots of different sizes have passed.
- g) If a shipment is found to fail test requirements it will be rejected by QC inspector and not released for shipment.

13. Packaging

100 pieces will be packed into a dispenser box. The number contained in each dispenser box determined by weight. 10 dispenser boxes are packed in 2 layers of 5 giving 10 dispenser boxes inside export quality shipper carton (outer).

14. Shipper Carton Specifications

SIZE	LENGTH (mm)	WIDTH (mm)	HEIGHT (mm)	GROSS WEIGHT (kg)
SMALL	310	250	232	4.0
MEDIUM	310	250	232	4.5
LARGE	310	250	232	5.0
XL	310	250	232	5.5

15. Storage Conditions

The gloves will maintain their properties when stored in dry conditions out of direct sunlight at between 10°C to 30°C.

16. Shelf Life

The gloves shall have a shelf life of 3 years from the date of manufacture when stored in compliance with the above storage conditions.